

Certificate of Testing

FT Technologies Limited

Church Lane Teddington Middlesex TW11 8PA

Serial Number: 12511CC44A

Page 1 of 2 Pages

Client's Order Number: P16265 Works Order Number: 12511-02

Date of Test: 28th April to 1st May 2010

Attn.: Mr. Olivier Hus

Specimens: 1 off FT702LT Wind Sensor

> Serial No.: 2870-001

Part No.: 22 (With Soldered Internal Connector)

TRaC Stores No.:

19th April 2010 Receipt Date:

Specification: Change of Temperature Test

Tested in accordance with BS EN 60068-2-14: 2009, Test N.

+70°C Upper Temperature: -40°C Lower Temperature:

Rate of Change: 5°C/minute

Duration: 2 hours at each extreme

Number of Cycles: 16

The specimen was attached to a support framework inside the test chamber and a **Procedure:**

platinum resistance thermometer was positioned adjacent to the test specimen to record the ambient air temperature during the test. A representative of the customer then connected the specimen under test to function test equipment located external to the test chamber. The specimen was placed in an operational state by the customer and its operation was continuously logged by the customer's laptop. The specimen was then tested in accordance with the specification and on completion of the test the specimen was visually inspected for any conspicuous signs of external damage or degradation.

O.A. APPROVAL D. C. Moni

D. Wheatley

D.K.Morris - Chief Test Engineer

Certfied that the specimens detailed hereon have been subjected to the tests as required by the order unless otherwise stated above. Our technical competence and quality control arrangements are in accordance with the conditions of our UKAS accreditation No representation or warranty is given that the Tests performed under the terms of the Contract constitute, in themsleves, a sufficient programme for the Customer's purpose, nor that the Customer's Equipment is suitable for any particular purpose The contents of this Certificate shall not be reproduced, except in full, without the written approval of TRaC Global Limited.



Rothwell Road, Warwick, CV34 5JX, UK. T +44 (0)1926 478478 F +44 (0)1926 478479 E test@tracglobal.com www.tracglobal.com

Issue Date: 7th July 2010





Certificate of Testing

Serial Number: 12511CC44A

Page 2 of 2 Pages

Results:

The specimen completed the test with no conspicuous signs of external damage or degradation and the customer reported that the specimen had operated correctly throughout the test.